

## FIELD SAFETY NOTICE: TRITEMP THERMOMETER (TR1)

FSN-2020-07-14

Henleys ref: RMA-001171

Henleys Medical Supplies is distributing this advisory field safety notice (FSN) on behalf of the manufacturer Trimedika. All relevant information is contained within the attached FSN.

This FSN relates to users unintentionally accessing the MODE settings and provides a revised Instructions For Use to ensure that if you inadvertently enter the MODE settings you will be aware of this action and will understand how to exit these settings.

Please ensure the new instructions for use are distributed to all end users.

**There is no requirement for Customers to return any TriTemp Thermometers to TriMedika as these devices can continue to be used in accordance with the updated IFU provided with this FSN.**

Please do not return any items to Henleys Medical Supplies under this FSN, but please do ensure the attached FSN Acknowledgment Form (page 5 of the attached FSN) is completed and returned to [vikki.patis@henleysmed.com](mailto:vikki.patis@henleysmed.com) as soon as possible.

If you have any questions, please get in touch using the details below and quoting RMA-001171.

Regards,



Vikki Patis  
Regulatory & Compliance Manager  
[vikki.patis@henleysmed.com](mailto:vikki.patis@henleysmed.com)  
01707 385223  
07788282609

