

To the attention of:

Pharmacist responsible for Sterile Medical Devices
and Local Medical Device Vigilance Associate

Lissieu, 13/07/2021

N/Réf : DA- R1910265 – RC19228/C

Objet : Field Safety Notice Enteral Feeding Tube

Urgent Field Safety Notice

Sir, Madam,

During our post-market surveillance process, we observed in some products a non-uniform yellow color on a portion of the feeding tube. In view of the increasing number of ppm observed and after discussion with ANSM, CAIR LGL has decided to proceed with product safety information on adult and paediatric enteral nutrition tubes (list of references in appendix A):

References : NCE...PU
NCE...AU
NCE...G

This phenomenon is characterized by the appearance of a non-uniform yellow color at different part on the polyurethane tubes of our adult and paediatric nutrition probes. We were aware that the natural oxidation of polyurethane could cause a slight yellow coloration of the polyurethane, but not in the proportions observed.



This defect is detectable as soon as the blister is opened and therefore systematically before placing the probe into the patient. All of our tests show that there is no accelerated degradation of the structure of the polyurethane, but in the absence of knowledge of the consequences on the patient, CAIR LGL recommends not to use feeding tubes with this coloration.

To correct this defect, an action has been taken internally and it is applicable as of now.

At the same time, we ask you to confirm receipt of this information by returning the form attached below.

As this warning follows a materiovigilance procedure, the competent authorities concerned are informed of this decision, and we will keep them informed of the progress of the operation.

We would like to thank you for your collaboration and through this warning, CAIR LGL reaffirms its desire to provide you with secure and efficient products.

Remaining at your entire disposal for any further information, please accept, Madam, Sir, the assurance of our best regards.

Delphine ALLARD
Correspondant Matériovigilance

Annex A :

Exhaustive list of references of enteral feeding tubes

NCE04PU	NCE110G	NCE214AU	NCE508G
NCE05NPU	NCE110NG	NCE214G	NCE508NAU
NCE05PU	NCE110PU	NCE306G	NCE508NG
NCE06NPU	NCE112AU	NCE306NPU	NCE510AU
NCE06PU	NCE112G	NCE306PU	NCE510G
NCE08NPU	NCE114AU	NCE308G	NCE512AU
NCE08PU	NCE114G	NCE308NG	NCE512G
NCE104PU	NCE114NAU	NCE308NPU	NCE514AU
NCE105PU	NCE116AU	NCE308PU	NCE514G
NCE106AU	NCE116G	NCE310G	NCE516AU
NCE106G	NCE118AU	NCE310NPU	NCE516G
NCE106NAU	NCE120AU	NCE310PU	NCE518AU
NCE106PU	NCE206AU	NCE312G	NCE520AU
NCE108AU	NCE206G	NCE312NG	NCE520NAU
NCE108G	NCE208AU	NCE312NPU	
NCE108NAU	NCE208G	NCE312PU	
NCE108NG	NCE210AU	NCE314G	
NCE108PU	NCE210G	NCE506AU	
NCE10PU	NCE212AU	NCE506G	
NCE110AU	NCE212G	NCE508AU	



Au commencement, il y avait un patient...

ACKNOWLEDGEMENT OF RECEIPT FORM SAFETY COMMUNICATION
FSN - DA- R1910265 – RC19228/C

To be completed and returned

Email: FieldAction@cairgl.fr

Fax: 0033 (0)4 78 43 77 07

Establishment (Stamp)

We hereby certify receipt of the safety communication / important notice sent by CAIR LGL on 21.07.2021 and its transmission to everyone needing to be informed of it within our organization.

We certify that we have read and understood the field safety notice

Name / Position of the person who completed the form

Date :

Signature :



ISO 9001 - ISO 13485

Service Commercial

Tél. +33 (0) 4 78 43 78 25

Fax +33 (0) 4 78 43 77 78

commercial@cairgl.fr

Service Commandes

Tél. +33 (0) 4 78 43 76 43

Fax +33 (0) 4 78 43 77 09

commande@cairgl.fr

Service Export

Tél. +33 (0) 4 78 43 77 44

Fax +33 (0) 4 78 43 77 09

export@cairgl.fr