

**URGENT MEDICAL DEVICE RECALL
RESPONSE CARD**

Device Name: injeTAK Adjustable Tip Needle

Model Numbers: DIS199 and DIS201

Lot Numbers: Reference Table 1 and Table 2

FACILITY INFORMATION	
Name of Facility	
Address of Facility	
Representative Name	
Representative Title	
Representative Email	

Does your facility have any units of DIS199 and /or DIS201 from any of the lots mentioned in **Table 1** and **Table 2**?

If you do **not** have any of the lots listed below, please check the box below to confirm that there are no affected devices at your facility:

I confirm that there are no affected devices from the above lot at the facility.

(If you have confirmed that there are no devices at your facility, please complete signature section on page 3)

If you do have any of the lots listed below, please follow the below instructions.

1. Customers are requested to return the product to the origin of their shipment. **Include this form with the shipment**, and please send it to:

US, Canada, and Australia Only Customers:

Laborie Medical Technologies Corp.
ATTN: Brad Schwarz
180 International Drive
Portsmouth, NH 03801 USA

Rest of World Customers:

Laborie Medical Technologies Corp.
ATTN: Receiving, injeTAK Returns
Colosseum 25, 7521 PV
Enschede The Netherlands

2. Please indicate what you have on hand and the quantity that will be returned in **Table 1** and **Table 2** below.

Table 1: DIS199 Lot Numbers

Lot Number	Quantity on Hand	Quantity Returned	Lot Number	Quantity on Hand	Quantity Returned
D198592			D211823		
D19C094			D211825		
D197765			D215573		
D19B105			D214446		
D19C097			D213099		
D19C742			D214448		
D201666			D213385		
D202622			D216690		
D203779			D217602		
D204597			D21A632		
D198638			D21B406		
D204100			D219650		
D201620			D221275		
D207081			D218636		
D205443			D21C155		
D205441			D221088		
D205445			D222083		
D207623			D223087		
D20A091			D224083		
D20C092			D224386		
D208579			D225586		
D20C094			D225589		
D20A092			D227088		
D20B085			D226108		
D211094			D228087		

See Table 2 on next page.

Table 2: DIS201 Lot Numbers

Lot Number	Quantity on Hand	Quantity Returned	Lot Number	Quantity on Hand	Quantity Returned
D19B109			D214447		
D199716			D211094		
D19B106			D214449		
D19C789			D215574		
D19C788			D216689		
D19C098			D219083		
D19C743			D217081		
D201667			D218089		
D203778			D21B407		
D202623			D21B081		
D204598			D21A082		
D205444			D221277		
D205440			D21C156		
D203783			D221089		
D205442			D221276		
D204101			D222084		
D203780			D224084		
D207624			D221087		
D207082			D222082		
D208580			D223088		
D205446			D226109		
D20B086			D224387		
D20A093			D227089		
D20C095			D228088		
D20C093			D226107		
D211824			D228197		
D211095			D229088		
D211826			D22A099		
D213100			D22A719		
D213386			D229086		

For the ease of processing, please provide Original P.O./Invoice #:

Representative Print Name

Representative Signature

Date

Please send completed and signed response card to LMTrecalls at LMTrecalls@laborie.com.