

26th May 2023

Dear Customer

Integra Product Recall

FSN Reference Number: FSN 2023-HHE-005

**SurgiMend® PRS, SurgiMend® PRS Meshed, SurgiMend®, SurgiMend® MP,
PriMatrix®, PriMatrix® Ag**

SurgiMend is to be removed from use immediately

We have been advised by our Supplier, Integra Lifesciences, that they have registered an FSCA (Field Safety Corrective Actions) with MHRA (Medicines and Healthcare products Regulatory Agency) and they are issuing a Field Safety Notice (FSN), which includes a full worldwide voluntary product recall of medical devices including SurgiMend.

The FSCA/FSN is following an internal investigation process that was conducted at the Integra Boston Facility whereby deviations during endotoxin testing were found, which may have resulted in the release of products with higher levels of endotoxins than permitted by the product specifications. Higher levels of endotoxins can induce an immune response, leading to a post-operative fever.

Our Administration team is currently liaising with all customers to ensure all devices are collected and returned to head office. Please use the contact details below if you require assistance or further information. Please note that all returned devices will be credited.

You will be sent a copy of the Field Safety Notice (FSN) issued by Integra, an MHRA explanatory flyer, and a copy of the customer response, which we require you to complete and return to us.



Q Medical Technologies Limited
Summerlands Trading Estate
Endmoor, Kendal, Cumbria, LA8 0FB
T: 0845 194 9284 F: 0845 094 9380
E: info@qmedical.co.uk W: www.qmedical.co.uk
Company Number: 5339771



Integra Lifesciences are working extremely hard to ensure that they can fulfill the requirements of the endotoxin testing and place their device back into the marketplace. Until such time, SurgiMend will not be available for purchase, and we will inform you of any changes to this situation as they arise.

We thank you for your support during this time.

Yours faithfully



Kristi Pillans MSc

Group Director for Quality & Regulatory Affairs

Person Responsible for Regulatory Compliance (PRRC)

Product Recall Contact Information

Q Medical Technologies Ltd

Unit 1a Summerlands Trading Estate

Endmoor

Kendal

Cumbria LA8 0FB

0845 1949284

Info@qmedical.co.uk