

### **Urgent: Customer Notice**

### TC-I Reusable Tip Clip Sensor (see Attachment 2 for Affected Part and Lot Numbers)

June 12, 2023

\*\*\*NHS Supply Chain\*\*\*
Accounts Payable
PO Box 253
Wymondham, Norfolk, NR18 8DL
United Kingdom

Attention: Clinical Engineering/Biomedical Department

Customer ID: NH0010\$E

### **Reason for Notice:**

Masimo identified TC- I Reusable Tip Clip sensors with higher-than-expected variability in SpO2 readings that are outside of the accuracy specification. Masimo's investigation found the TC-I sensors had higher variability with an overall Arms of 4.4% compared to the specification of 3.5% Arms.

The Customer Notice applies to the specific TC-I Reusable Tip Clip Sensor with the part and lot number(s) identified in this communication. Please assess your inventory and promptly remove all sensors with part numbers and lot numbers listed below from use and inventory.

Please refer to Attachment 2 for a listing of the TC-I Reusable Tip Clip Sensor(s) shipped to \*\*\*NHS Supply Chain\*\*\* subject to this Customer Notice.

#### Risk to Health:

Inaccurate readings could potentially result in a delay in care or unnecessary medical intervention. Masimo has not received any complaints alleging adverse events.

#### No Impact to Other Products:

There is no impact to any other Masimo products.

### Actions to be taken by Customer:

Contact Masimo Technical Services to request the return of the affected part and lot number(s) for replacements and/or if you have part numbers or lot numbers not included in this communication and have questions about those parts. Masimo's Technical Services is available Monday through Friday 6:00am - 5:00pm PST (BST -8) for live assistance or 24 hours a day to leave a message and someone will contact you. Local contact information can be found at <a href="http://service.masimo.com">http://service.masimo.com</a>. Complete Attachment 1, then email it to CustomerNotice@masimo.com, or Fax it to 1-949-297-7575 by 12 July 2023.

Masimo is committed to consistently providing high quality products and services to you, our customers. We apologise for any inconvenience this issue may have caused.

Sincerely,

Mathew Jimenez

SVP, Worldwide Quality and Compliance Masimo Corporation



# **Attachment 1**

## Action to be taken:

1. Please follow the steps below .

Contact Masimo Technical Services to request the return of the affected part and lot number(s) to Masimo for replacement and indicate which, if any, of the affected part and lot number(s) is (are) no longer in service. If you have part numbers or lot numbers not included in this communication and have questions about those parts, please contact Masimo Technical Services. Local contact information can be found at <a href="http://service.masimo.com">http://service.masimo.com</a>.

Complete the bottom portion of this *Attachment*, then email it to <u>CustomerNotice@masimo.com</u> or Fax it to 1-949-297-7575 <u>by 12 July 2023.</u>

. On behalf of my facility, I a one.	acknowledge receipt of the Customer notice titled TC-	I Reusable Tip Clip Sensor. Select
<del></del>	tacted Masimo Technical Services and arranged for table (s) to Masimo for replacement.	the return of the affected part
b. None of the at my facilit	e potentially affected part and lot number(s) listed in $\lambda$ ty.	Attachment 2 is (are) in service
. Sign below, then Email th July 2023.	nis attachment to <u>CustomerNotice@masimo.com</u> , or	Fax it to 1-949-297-7575 <b>by 12</b>
Contact Information and Authorisation	Customer Name	Customer ID
	***NHS Supply Chain***	NH0010\$E
$\rightarrow$		
→ Printed Name	Signature	Date



# **Attachment 2**

Part and lot number(s) shipped to your facility:

Part number: 1895, Lot number: 17J09, 17MDD, 17N06, 18BAA, 18CCT, 18CWV, 18CYG, 18DCB, 18EEC, 18ENJ,

18FAG, 18G1V, 18GND, 18H3N, 18K47, 19AAA, 19AHL, 19BFC, 19CAC

Part number: 2503, Lot number: 17J12, 17MDF, 17N09, 18AVB, 18FQZ