

**Our Ref:** DCCV-NC03422

1<sup>st</sup> March 2023

Dear Customer,

Please find attached MEDICAL DEVICE PRODUCT CORRECTION letter, issued in relation to product notification involving the Air-Q3 models with gastric access.

Fannin Ltd. has been informed that the labeling in the instructions for use (IFU) and the printed text on the device indicate an orogastric (OG) tube size discrepancy.

Please refer to correct recommended maximum OG Tube/Catheter sizes for each Air-Q3 model listed in the supplier's letter attached.

The product has been evaluated by the manufacturer as safe to use and not considered high risk.

Please note this is not a Field Safety Notice - any products that you might still have in stock do not need to be returned and can be used, taking under consideration the correct information listed in the attached letter.

Please complete the Product Notification Response Form and return at your earliest convenience to [Louise.ford@fannin.eu](mailto:Louise.ford@fannin.eu). Alternatively, please reply to your email, stating that you have read and understood the product correction letter.

Kind Regards

**Joanna Rakowska**

Quality & Regulatory Specialist  
E: [joanna.rakowska@dccvital.com](mailto:joanna.rakowska@dccvital.com)

**DCC VITAL**

South County Business Park  
Leopardstown  
Dublin 18  
Ireland

T: +353 (0) 1 290 7000  
F: +353 (0) 1 290 7119

[www.DCCVital.com](http://www.DCCVital.com)