

3M Company

3M Health Care  
2510 Conway Ave., Bldg. 275-5W-06  
St. Paul, MN 55144 U.S.A.

February 16, 2024



**Incorrect Dressing Dimension on Primary Carton of  
3M™ Tegaderm™ CHG Chlorhexidine Gluconate I.V. Securement Dressing (1657R)**

Select lots of the 3M™ Tegaderm™ CHG Chlorhexidine Gluconate I.V. Securement Dressing (catalog # 1657R, lot numbers listed below) have an incorrect dressing dimension stated on the primary carton. The dimension is reported as 8.5 cm x 8.5 cm and the correct dimension is 8.5 cm x 11.5 cm. The error affects only the metric dimension. The imperial dimension (inches) stated is correct. The dimensions are correctly listed on the shipper and pouch level and, therefore, are available at the point of use. There is no impact to fit, form, safety, or performance of the product. This is solely a labeling error.

**Affected lot numbers (production period September 2023 to January 2024):**

33T7CN, 33TWLT, 33TX4T, 33TXLM, 33TXWR, 33TY6P, 33TYCN, 33TYKT, 33TYTE,  
33W36J, 33W3H7, 33W3RC, 33W44X, 33W4JF, 33W4TC, 33WA8R, 33WAKY, 33WC65,  
33WCFM, 33WCTN, 33WDDW, 33WD87, 33WC65, 33WETX, 33WEET, 33WE94,  
33WFD9

3M has corrected this labeling error.

Sincerely,

Nadia Battah  
Regulatory Affairs Manager  
Medical Surgical Business  
3M Company