



Creo Medical UK Ltd  
Bridgend Road Industrial Estate,  
Dingwall, Ross-shire  
IV15 9QF  
Scotland UK

Tel: +01904 786888  
Fax: +01904 787818  
E-mail: [customerservice@creomedical.com](mailto:customerservice@creomedical.com)  
Website: [www.creomedical.com](http://www.creomedical.com)

To whom it may concern,

In relation to reported faults of CPS-2410235336 in which the snare loop has split or broken to date affected Lots reported are

ME250729M303

ME251108M304

ME251122M302

ME251105M303

ME251122M303

ME250901M301

ME251105M305

ME250727M302

In relation to clarification documents from the manufacturer there were 4 article numbers potentially affected however only one of these additional article numbers has been supplied in the UK – CPS-2415235336. This is a much lower volume product and we do not believe any of the faulty alternative have been supplied.

With respect to any orders placed through Supply Chain since the issue was raised we have only been supplying the original model (manufacturer AGS) of CPS-2410235336 as per the framework. We have had further deliveries of stock this week and additional deliveries due in early July. We have also begun to receive new batches of stock from new manufacturer (Transmed) for which the issue has been corrected. Submission for this version of the product to be placed on Supply Chain has already been made. Corrective actions on the improvement made to the product along with FSN has been provided in a separate document.

Kind Regards,

A handwritten signature in black ink, appearing to read "Neil Bottomley".

Neil Bottomley  
Endoscopy Sales Manager  
Creo Medical UK Ltd

A handwritten signature in black ink, appearing to read "Luis Collantes".

Luis Collantes  
CEO  
Creo Medical UK Ltd

**Anything is possible with the right approach**