

29 July 2026

Subject: *Field Safety Notice (FSN RKL-7079) – SKINTACT Defibrillation Electrodes*
FGB-FSN-00003 (FSN-RKL-7079)

Dear Customer,

Please find attached a Field Safety Notice (FSN Ref: RKL-7079) issued by the legal manufacturer, Leonhard Lang GmbH, relating to specific lots of the following SKINTACT defibrillation electrodes:

- REF DF28N – SKINTACT Adult Electrodes
- REF DF30G – SKINTACT Paediatric Electrodes

Reason for Field Safety Notice:

The connector cannot be fully inserted into the compatible adapter cable using the usual amount of force. The locking tab does not engage and the audible “click” is absent.

Importantly, the manufacturer's investigation has confirmed that:

“Electrodes can be used without restrictions”

Actions required:

- Review the attached Field Safety Notice (FSN RKL-7079).
- Inform relevant personnel within your organisation as appropriate.
- Continue using affected products as normal; no restrictions on use apply.

This FSN is for awareness, no other actions are required.

Should you have any questions regarding this communication, please refer to the attached Field Safety Notice or contact your Fannin representative.

Yours sincerely,

Nischith Raphael
Head QA&RA
Fannin UK Ltd