B. Braun Medical Ltd
Thorncliffe Park
Sheffield
S35 2PW
Tel: 0114 225 9000
Fax: 0114 225 9111

04 Nov 2026

Infusomat® SafeSet AirStop Filter Specification Update

Dear Customer,

We would like to inform you of an update to the declared filter specification within the Infusomat® SafeSet AirStop administration set portfolio.

Following Medical Device Regulation (MDR) revalidation using current ISO test methodologies, the declared filter **specification** has been updated from 15 µm to 3 µm.

This update reflects the application of current regulatory requirements and test methodologies during MDR revalidation. **The AirStop membrane has not changed. Its material properties, membrane structure, manufacturing process, intended purpose and filtration performance remain unchanged.**

Product packaging bearing the revised 3 µm declared filter specification have been manufactured for global supply since July 2026 and are now entering the UK market as existing stock is depleted.

For customers requiring alternative filtration specifications, suitable options remain available within the B. Braun portfolio, including:

- Administration sets incorporating a 15 µm filter
- Transfusion administration sets incorporating a 200 µm filter
- In-line filters available with 1.2 µm and 0.2 µm filtration options.

As with all intravenous administration systems, healthcare professionals should review the relevant Summary of Product Characteristics (SmPC) and local policies to confirm compatibility between the selected administration device and the medicinal product being administered.

Should you have any questions regarding this communication, please contact your local B. Braun representative or B. Braun Medical Information: medinfo.bbmuk@bbraun.com or 0800 298 0299.

Kind regards,

B. Braun Medical UK